

# EC Declaration of Conformity

<b>Manufacturer/ Supplier Information:</b>	<b>BioFire Diagnostics, LLC</b> 515 Colorow Drive Salt Lake City, Utah 84108, USA Phone: 1-801-736-6354 <a href="mailto:regulatory@BioFireDX.com">regulatory@BioFireDX.com</a> <a href="http://www.BioFireDX.com">http://www.BioFireDX.com</a>
--	---

We, BioFire Diagnostics, LLC, declare under our sole responsibility, that the product

## FilmArray 2.0 Instrument (FLM2-ASY-0001)

meets the provisions of the European Directive 98/79/EC for *in vitro* Diagnostic Medical Devices, the European Directive 2011/65/EU on the restriction of the use of certain hazardous substances (ROHS) in electrical and electronic equipment, and the European Directive 2012/19/EU on waste electrical and electronic equipment (WEEE). The device is classified as a General In Vitro Diagnostic (IVD) device.

BioFire Diagnostics' quality system is registered to ISO 13485:2003 and EN ISO 13485:2012.

The following relevant standards have been met:

<b>ISO 13485:2003/EN ISO 13485:2012</b> Medical devices – Quality Management System – Requirements for regulatory purposes
<b>EN ISO 14971:2012</b> Medical devices – Application of risk management to medical devices
<b>EN 62366:2008</b> Medical devices-Application of usability engineering to medical devices
<b>EN 13612:2002</b> Performance evaluation of in vitro diagnostic devices
<b>EN 61010-2-101:2002</b> Safety requirements for electrical equipment for measurement, control, and laboratory use --Part 2.
<b>EN 61326-2-6:2006</b> Electrical equipment for measurement, control, and laboratory use - EMC requirements -- Part 2-6
<b>EN 62304:2006</b> Medical device software—Software life-cycle processes—IEC 62304:2006, November 27, 2008.
<b>EN 980:2008</b> Symbols for use in the labelling of medical devices
<b>ISO 15223-1:2012</b> Medical Devices – Symbols to be used with medical device labels, labeling and information to be supplied – Part 1: General requirements
<b>EN ISO 18113-1:2011</b> In vitro diagnostic medical devices – Information supplied by the manufacturer (labeling) – Part 1: Terms, definition and general requirements
<b>EN ISO 18113-3:2011</b> In vitro diagnostic medical devices – Information supplied by the manufacturer (labeling) – Part 3: In vitro diagnostic instruments for professional use

Technical documentation demonstrating compliance as described in Annex III of the European Directive 98/79/EC is kept by the manufacturer and can be made available by the authorized representative in Europe - QARAD bvba, Ciplastraat 3, B-2440 Geel, Belgium.

The notified body for this product is BSI (Notified body #0086; Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes, MK5 9PP, United Kingdom).

Salt Lake City, UT, USA      5/30/2017  
(Place and date of issue)

  
**Randy Rasmussen**  
President and Chief Executive Officer

**BIO FIRE** 515 Colorow Drive, Salt Lake City, UT 84108  
phone 1-801-736-6354 | fax 1-801-588-0507  
www.BioFireDX.com  
A BIOMÉRIEUX COMPANY